

**J64030A (30A) TRANSMISSION MEASURING SET**  
**(RANGE: UP TO 150 KILOCYCLES)**

**DESCRIPTION**

**1. GENERAL**

(a) 4.07.1—added paragraph

**1.001** This addendum supplements Section 103-408-100, Issue 2.

**1.002** This addendum is issued to add a new paragraph concerning the use of the J94002AA milliwatt reference set.

**4.07.1** After the 30A transmission measuring set has been carefully calibrated, the deviation from a true reading should be measured by using a J94002AA (2AA) milliwatt reference set (if available) to check the transmission measuring set. The complete method of making this check is specified in Section 103-429-100.

The following change applies to Part 4 of the section: