

911A, B, AND C DATA TEST SETS DESCRIPTION, OPERATION, AND MAINTENANCE

1. GENERAL

1.001 This addendum supplements Section 103-813-100, Issue 3.

1.002 This addendum is issued to reflect the filter response time of the new filter provided by options A and ZB.

3. OPERATION

The following change applies to Part 3 of the section.

(a) Table D—revised

For FILTER switch, change value in FUNCTION column from "1.5 msec" to "100 μ sec."