

TRIPLETT MODEL 310C VOLT-OHM-MILLIAMMETER

1. GENERAL

1.01 This addendum is issued to correct meter range information in paragraph 3.02, c.

1.02 With red pencil or ink, make the changes specified in paragraph 2 of this addendum, and file the addendum in front of CTSP 405-110-301.

2. CORRECTION

2.01 Change paragraph 3.02, c, to read as follows:

c. The 0-600 range is read on the 0-60 scale by adding one zero to the reading; e.g., a meter reading of 40 would equal 400 volts.