



iHealth Edge Battery Replacement

In this guide we will show how to replace the battery in the iHealth Edge.

Written By: Gabrielle Bettegnies



INTRODUCTION

This guide will show you how to replace the battery on the iHealth Edge. You must take the device and remove the two piece that are stuck together using a blow dryer. This will expose the inside of the device to see screen. This is where we will begin to learn how to replace the battery.

TOOLS:


- [Hair Dryer](#) (1)
- [iFixit Opening Tools](#) (1)
- [Soldering Iron](#) (1)

Step 1 — Battery



- Lightly heat device with hair dryer for 15 seconds to soften adhesive.

Step 2




- After heating adhesive, firmly pry apart the two halves.

Step 3




- Remove polymer cover that is protecting the mother board, it should just come out when two halves are separated.

Step 4




- Use a spudger or plastic opening tool to assist the removal of the motherboard. Pry up on two tabs and the mother board will pop out.

Step 5



- This shows the mother board and battery out and exposed.

Step 6



- You must remove the solder from two wires from the motherboard that are connected to the battery.

To reassemble your device, follow these instructions in reverse order.